LABORATORY REPORT

Optimal Balance Pharmacy

,

Phone: 281-919-2964

Fax: 346-270-2938

Account #: E19193
Sample Name: ARA-290
Lot #: LG342010854
Storage Conditions: Refrigerated

Date Received: 9/23/2026

Submission #: ETX-260923-0371

Microbiology Tests:

Celsis Sterility Test Date Started: 9/23/2026 Date Completed: 10/1/2026

Method Suitability: 2510290110
Method Suitability - Technique: Direct Inoculation

Test Record:	100126-2011
Method Performed:	Direct Inoculation
Positive Cut-off RLU:	88268
TSB RLU:	847
FTM RLU:	1434
Day 7 Result:	Pass

USP <85> Bacterial Endotoxin Test Date Started: 9/24/2026 Date Completed: 9/24/2026

Numerical Value :	<2 EU/mL
Endotoxin Limit:	500 EU/mL
Result:	Pass

Respectfully submitted,
EAGLE ANALYTICAL SERVICES INC.

Alison Eng, Quality Control Manager

Results apply only to the tested sample(s) as received.